

CM05-M01 - Dec. 1, 2016

Item # CM05-M01 was discontinued on Dec. 1, 2016. For informational purposes, this is a copy of the website content at that time and is valid only for the stated product.

16 MM CAGE CUBE-MOUNTED TURNING PRISM MIRRORS

- ▶ **Metallic- or Dielectric-Coated Turning Prism Mirrors**
- ▶ **Premounted in 16 mm Cage Cubes**
- ▶ **Compatible with SM05 Lens Tubes and 16 mm Cage Systems**



CM05-G01



4-40 Tapped Holes Provide
Compatibility with Thorlabs'
16 mm Cage System

OVERVIEW**Features**

- Choose from Seven Coating Options
 - Metallic: UV-Enhanced Aluminum, Protected Aluminum, Protected Gold, Protected Silver
 - Dielectric: E02 (400 - 750 nm), E03 (750 - 1100 nm), or K13 Nd:YAG (532 nm and 1064 nm)
- Compatible with Our Ø1/2" Lens Tubes and 16 mm Cage System
- Part Number and Coating are Engraved on the Housing for Easy Identification
- Post Mountable



Click to Enlarge
Cage Cube Shown with SM05 Lens Tubes

These Cage-Cube-Mounted Turning Mirrors provide additional flexibility when building optical devices based upon our 16 mm cage system and SM05 lens tube products. The mounted turning mirrors are prealigned so that the reflected beam exits to within ± 20 arcmin of 90° , which can reduce the time needed for aligning the mirror in a setup. For higher accuracy or more alignment flexibility, see our 90° turning mirror kinematic mounts. These turning mirrors are available with any of seven reflective coatings: UV-enhanced aluminum, protected aluminum, protected silver, protected gold, E02 broadband dielectric (400 - 750 nm), E03 broadband dielectric (750 - 1100 nm), or K13 Nd:YAG dielectric (532 nm and 1064 nm). A complete list of mirror specifications is provided on the *Specs* tab. Reflectance plots and data are provided on the *Graphs* tab.

The bottom of the CM05 cubes are M6 x 0.5 threaded, and 8-32 and M4 adapters are included for post mounting, while the CCM5 cubes have an M4 tapped hole on the bottom. The entrance and exit ports of both types of cubes have SM05 threading (0.535"-40) as well as 4-40 taps that accept our SM05 Lens Tubes and Ø4 mm cage rods, respectively (refer to the images to the right and above). These mounted turning mirrors can be connected to other cage cubes through the use of our cage rods and SRSCA adapters.

Item #	CM05-F01 and CCM5-F01/M	CM05-G01 and CCM5-G01/M	CM05-P01 and CCM5-P01/M	CM05-M01 and CCM5-M01/M	CM05-E02 and CCM5-E02/M	CM05-E03 and CCM5-E03/M	CM05-K13 and CCM5-K13/M
Unmounted Prism Item #	MRA20-F01	MRA20-G01	MRA20-P01	MRA20-M01	MRA20-E02	MRA20-E03	MRA20-K13
Reflective Coating	UV-Enhanced Aluminum	Protected Aluminum	Protected Silver	Protected Gold	E02 Dielectric	E03 Dielectric	Nd:YAG Dielectric
Reflectivity for 0° to 45° AOI	$R_{avg} > 85\%$ (250 - 600 nm)	$R_{avg} > 90\%$ (400 nm - 10 μm)	$R_{avg} > 96\%$ (400 - 700 nm) $R_{avg} > 97.5\%$ (700 nm - 2.0 μm)	$R_{avg} > 96\%$ (800 nm - 20.0 μm)	$R_{avg} > 99\%$ (400 - 750 nm)	$R_{avg} > 99\%$ (750 - 1100 nm)	R_s and R_p : >98% (532 nm) >99% (1064 nm)

S P E C S

Metallic-Coated Mirrors

Item #	CM05-F01 and CCM5-F01/M	CM05-G01 and CCM5-G01/M	CM05-P01 and CCM5-P01/M	CM05-M01 and CCM5-M01/M
Unmounted Prism Item #	MRA20-F01	MRA20-G01	MRA20-P01	MRA20-M01
Reflective Coating	UV-Enhanced Aluminum	Protected Aluminum	Protected Silver	Protected Gold
Reflectivity (Wavelength Range)	$R_{avg} > 90\%$ (250 - 450 nm) ^a	$R_{avg} > 90\%$ (450 nm - 2 μm) $R_{avg} > 95\%$ (2 - 20 μm) ^b	$R_{avg} > 96\%$ (2.0 - 20.0 μm) $R_{avg} > 97.5\%$ (450 nm - 2.0 μm) ^c	$R_{avg} > 96\%$ (800 nm - 20.0 μm)
Damage Threshold (10 ns pulse, 10 Hz)	0.3 J/cm ² (355 nm, Ø0.381 mm)	0.3 J/cm ² (1064 nm, Ø1.000 mm)	3 J/cm ² (1064 nm, Ø1.000 mm)	2 J/cm ² (1064 nm, Ø1.000 mm)
CW Damage Threshold ^d	300 W/cm (1.064 μm , Ø0.044 mm) 500 W/cm (10.6 μm , Ø0.339 mm)	60 W/cm (1.064 μm , Ø0.044 mm) 350 W/cm (10.6 μm , Ø0.339 mm)	1750 W/cm (1.064 μm , Ø0.044 mm) 1500 W/cm (10.6 μm , Ø0.339 mm)	1500 W/cm (1.064 μm , Ø0.044 mm) 750 W/cm (10.6 μm , Ø0.339 mm)
Ports	2 Ports, Each with SM05 (0.535"-40) Threading and Four 4-40 Taps for Cage Rods			
Post Mounting	CM05 Prefix: M6-Tapped Hole with 8-32 and M4 Adapters CCM5 Prefix: M4-Tapped Hole			
Mirror Substrate	N-BK7			
Housing Material	Engraved Black Anodized Aluminum Housing			
Surface Flatness (@ 633 nm)	$\lambda/10$ (Over Clear Aperture)			
Clear Aperture	Ø12.5 mm			
Surface Quality	40-20 Scratch-Dig			
Beam Path Deviation	90° ± 20 arcmin			


- For CM05-F01, $R_{avg} > 85\%$ (250 - 600 nm).
- For CM05-G01, $R_{avg} > 90\%$ (400 nm - 10 μm).
- For CM05-P01, $R_{avg} > 96\%$ (400 - 700 nm) and $R_{avg} > 97.5\%$ (700 nm - 2.0 μm).
- The power density of your beam should be calculated in terms of W/cm. For an explanation of why the linear power density provides the best metric for long pulse and CW sources, please see the *Damage Thresholds* tab.

Dielectric-Coated Mirrors

Item #	CM05-E02 and CCM5-E02/M	CM05-E03 and CCM5-E03/M	CM05-K13 and CCM5-K13/M
Unmounted Prism Item #	MRA20-E02	MRA20-E03	MRA20-K13
Reflective Coating	E02 Dielectric	E03 Dielectric	Nd:YAG Dielectric


Reflectivity (Wavelength Range)	$R_{avg} > 99\%$ (400 - 750 nm)	$R_{avg} > 99\%$ (750 - 1100 nm)	R_s and R_p : >98% (532 nm) >99% (1064 nm)
Damage Threshold (10 ns pulse, 10 Hz)	0.25 J/cm ² (532 nm, Ø0.803 mm)	1.0 J/cm ² (810 nm, Ø0.133 mm) 0.5 J/cm ² (1064 nm, Ø0.433 mm)	8 J/cm ² (532 nm, Ø0.491 mm) 5 J/cm ² (1064 nm, Ø1.010 mm)
Ports	2 Ports, Each with SM05 (0.535"-40) Threading and Four 4-40 Taps for Cage Rods		
Post Mounting	CM05 Prefix: M6-Tapped Hole with 8-32 and M4 Adapters CCM5 Prefix: M4-Tapped Hole		
Mirror Substrate	N-BK7		
Housing Material	Engraved Black Anodized Aluminum Housing		
Surface Flatness (@ 633 nm)	UV-Enhanced Aluminum Coating (250 - 450 nm) λ/10 (Over Clear Aperture)		
Clear Aperture	Ø12.5 mm		
Surface Quality	10-5 Scratch-Dig		
Beam Path Deviation	90° ± 20 arcmin		

The shaded regions in the graphs denote the ranges over which we guarantee the specified reflectance. Please note that the reflectance outside of these bands is typical and can vary from lot to lot, especially in out-of-band regions where the reflectance is fluctuating or sloped.



Click to Enlarge


Excel Spreadsheet with Raw Data for UV-Enhanced Aluminum, 12° and 45° AOI



Click to Enlarge


Excel Spreadsheet with Raw Data for UV-Enhanced Aluminum, 12° and 45° AOI

Protected Aluminum Coating (450 nm - 20 µm)




Click to Enlarge

Excel Spreadsheet with Raw Data for Protected Aluminum, 12° and 45° AOI



[Click to Enlarge](#)

[Excel Spreadsheet with Raw Data for Protected Gold, 12° and 45° AOI](#)




[Click to Enlarge](#)

[Excel Spreadsheet with Raw Data for Protected Silver, 12° and 45° AOI](#)

These plots show the reflectivity of our -E02 (400 - 750 nm) and -E03 (750 - 1100 nm) dielectric coatings for a typical coating run. The shaded region in each graph denotes the spectral range over which the coating is highly reflective. Due to variations in each run, this recommended spectral range is narrower than the actual range over which the optic will be highly reflective. If you have any concerns about the interpretation of this data, please contact Tech Support. For applications that require a mirror that bridges the spectral range between the dielectric coatings, please consider a metallic mirror.


-E02 Coating (400 - 750 nm)



[Click to Enlarge](#)

[Excel Spreadsheet with Raw Data for -E02 Coating, 8° and 45° AOI](#)

-E03 Coating (750 - 1100 nm)




[Click to Enlarge](#)


[Excel Spreadsheet with Raw Data for -E03 Coating, 8° and 45° AOI](#)

-K13 Coating (532 nm and 1064 nm)

These plots show the reflectance of our Nd:YAG (532 nm and 1064 nm) dielectric coating for a typical coating run. Although there will be variations in the broadband reflectance from run to run, this optic is guaranteed to meet the reflectivity specification at 532 nm and 1064 nm (see the table on the [Overview](#) tab). If you have any concerns about the interpretation of this data, please contact Tech Support.



Click to Enlarge











Click to Enlarge

Excel Spreadsheet with Raw Data for Dielectric Nd:YAG Laser Line Coating, 8° (Unpolarized) and 45° (S- and P-Polarized) AOI

ND:YAG OPTICS

Thorlabs offers a wide selection of optics optimized for use with Nd:YAG lasers. Please see below for more information.

Nd:YAG Optics Selection				
Dielectric Mirrors			Beamsplitters	
				
Laser Line Mirrors, 1064 nm, 532 nm, 355 nm, 266 nm	Right-Angle Prism Mirrors, 1064 nm, 532 nm	Cage Cube-Mounted Prism Mirrors, 1064 nm, 532 nm	Harmonic Beamsplitters, 1064 nm, 532 nm, 355 nm, 266 nm	High-Power Polarizing Beamsplitter Cubes, 1064 nm, 532 nm: Unmounted or Mounted
Lenses		Objectives	Filters	
				
UVFS Plano-Convex Lenses, 1064 nm, 532 nm: Unmounted or Mounted	Air-Spaced Doublets, 1064 nm, 532 nm	High Power Focusing Objectives, 1064 nm, 532 nm	Laser Line Filters, 1064 nm: Standard or Premium	Laser Line Filters, 532 nm: Standard or Premium

DAMAGE THRESHOLDS

Damage Threshold Data for Thorlabs' Turning Mirror Prisms

The specifications to the right are measured data for Thorlabs' turning mirror prisms.

Damage Threshold Specifications		
Item #	Type	Damage Threshold
CM05-F01 CCM5-F01/M	Pulse	0.3 J/cm ² (355 nm, 10 ns, 10 Hz, Ø0.381 mm)
	CW ^a	300 W/cm (1.064 µm, Ø0.044 mm) 500 W/cm (10.6 µm, Ø0.339 mm)
CM05-G01 CCM5-G01/M	Pulse	0.3 J/cm ² (1064 nm, 10 ns, 10 Hz, Ø1.000 mm)
	CW ^a	60 W/cm (1.064 µm, Ø0.044 mm) 350 W/cm (10.6 µm, Ø0.339 mm)
CM05-P01 CCM5-P01/M	Pulse	3 J/cm ² (1064 nm, 10 ns, 10 Hz, Ø1.000 mm)
	CW ^a	1750 W/cm (1.064 µm, Ø0.044 mm) 1500 W/cm (10.6 µm, Ø0.339 mm)
CM05-M01	Pulse	2 J/cm ² (1064 nm, 10 ns, 10 Hz, Ø1.000 mm)

CCM5-M01/M	CW ^a	1500 W/cm (1.064 μm, Ø0.044 mm) 750 W/cm (10.6 μm, Ø0.339 mm)
CM05-E02 CCM5-E02/M	Pulse	0.25 J/cm ² (532 nm, 10 ns, 10 Hz, Ø0.803 mm)
CM05-E03 CCM5-E03/M	Pulse	1.0 J/cm ² (810 nm, 10 ns, 10 Hz, Ø0.133 mm) 0.5 J/cm ² (1064 nm, 10 ns, 10 Hz, Ø0.433 mm)
CM05-K13 CCM5-K13/M	Pulse	8 J/cm ² (532 nm, 10 ns, 10 Hz, Ø0.491 mm) 5 J/cm ² (1064 nm, 10 ns, 10 Hz, Ø1.010 mm)

- The power density of your beam should be calculated in terms of W/cm. For an explanation of why the linear power density provides the best metric for long pulse and CW sources, please see the "Continuous Wave and Long-Pulse Lasers" section below.

Laser Induced Damage Threshold Tutorial

The following is a general overview of how laser induced damage thresholds are measured and how the values may be utilized in determining the appropriateness of an optic for a given application. When choosing optics, it is important to understand the Laser Induced Damage Threshold (LIDT) of the optics being used. The LIDT for an optic greatly depends on the type of laser you are using. Continuous wave (CW) lasers typically cause damage from thermal effects (absorption either in the coating or in the substrate). Pulsed lasers, on the other hand, often strip electrons from the lattice structure of an optic before causing thermal damage. Note that the guideline presented here assumes room temperature operation and optics in new condition (i.e., within scratch-dig spec, surface free of contamination, etc.). Because dust or other particles on the surface of an optic can cause damage at lower thresholds, we recommend keeping surfaces clean and free of debris. For more information on cleaning optics, please see our *Optics Cleaning* tutorial.

Testing Method


Thorlabs' LIDT testing is done in compliance with ISO/DIS11254 specifications. A standard 1-on-1 testing regime is performed to test the damage threshold.

First, a low-power/energy beam is directed to the optic under test. The optic is exposed in 10 locations to this laser beam for a set duration of time (CW) or number of pulses (pulse repetition frequency specified). After exposure, the optic is examined by a microscope (~100X magnification) for any visible damage. The number of locations that are damaged at a particular power/energy level is recorded. Next, the power/energy is either increased or decreased and the optic is exposed at 10 new locations. This process is repeated until damage is observed. The damage threshold is then assigned to be the highest power/energy that the optic can withstand without causing damage. A histogram such as that below represents the testing of one BB1-E02 mirror.



The photograph above is a protected aluminum-coated mirror after LIDT testing. In this particular test, it handled 0.43 J/cm² (1064 nm, 10 ns pulse, 10 Hz, Ø1.000 mm) before damage.

According to the test, the damage threshold of the mirror was 2.00 J/cm² (532 nm, 10 ns pulse, 10 Hz, Ø0.803 mm). Please keep in mind that these tests are performed on clean optics, as dirt and contamination can significantly lower the damage threshold of a component. While the test results are only representative of one coating run, Thorlabs specifies damage threshold values that account for coating variances.



Example Test Data			
Fluence	# of Tested Locations	Locations with Damage	Locations Without Damage
1.50 J/cm ²	10	0	10
1.75 J/cm ²	10	0	10
2.00 J/cm ²	10	0	10
2.25 J/cm ²	10	1	9
3.00 J/cm ²	10	1	9

Continuous Wave and Long-Pulse Lasers

5.00 J/cm ²	10	9	1
------------------------	----	---	---

When an optic is damaged by a continuous wave (CW) laser, it is usually due to the melting of the surface as a result of absorbing the laser's energy or damage to the optical coating (antireflection) [1]. Pulsed lasers with pulse lengths longer than 1 μs can be treated as CW lasers for LIDT discussions. Additionally, when pulse lengths are between 1 ns and 1 μs, LIDT can occur either because of absorption or a dielectric breakdown (must check both CW and pulsed LIDT). Absorption is either due to an intrinsic property of the optic or due to surface irregularities; thus LIDT values are only valid for optics meeting or exceeding the surface quality specifications given by a manufacturer. While many optics can handle high power CW lasers, cemented (e.g., achromatic doublets) or highly absorptive (e.g., ND filters) optics tend to have lower CW damage thresholds. These lower thresholds are due to absorption or scattering in the cement or metal coating.


Pulsed lasers with high pulse repetition frequencies (PRF) may behave similarly to CW beams. Unfortunately, this is highly dependent on factors such as absorption and thermal diffusivity, so there is no reliable method for determining when a high PRF laser will damage an optic due to thermal effects. For beams with a large PRF both the average and peak powers must be compared to the equivalent CW power. Additionally, for highly transparent materials, there is little to no drop in the LIDT with increasing PRF.

In order to use the specified CW damage threshold of an optic, it is necessary to know the following:


1. Wavelength of your laser
2. Linear power density of your beam (total power divided by 1/e² beam diameter)
3. Beam diameter of your beam (1/e²)
4. Approximate intensity profile of your beam (e.g., Gaussian)

The power density of your beam should be calculated in terms of W/cm. The graph to the right shows why expressing the LIDT as a linear power density provides the best metric for long pulse and CW sources. In this regime, the LIDT given as a linear power density can be applied to any beam diameter; one does not need to compute an adjusted LIDT to adjust for changes in spot size. This calculation assumes a uniform beam intensity profile. You must now consider hotspots in the beam or other non-uniform intensity profiles and roughly calculate a maximum power density. For reference, a Gaussian beam typically has a maximum power density that is twice that of the uniform beam (see lower right).

Now compare the maximum power density to that which is specified as the LIDT for the optic. If the optic was tested at a wavelength other than your operating wavelength, the damage threshold must be scaled appropriately. A good rule of thumb is that the damage threshold has a linear relationship with wavelength such that as you move to shorter wavelengths, the damage threshold decreases (i.e., a LIDT of 10 W/cm at 1310 nm scales to 5 W/cm at 655 nm):



LIDT in linear power density vs. pulse length and spot size. For long pulses to CW, linear power density becomes a constant with spot size. This graph was obtained from [1].



$$\text{Adjusted LIDT} = \text{LIDT Power} \left(\frac{\text{Your Wavelength}}{\text{LIDT Wavelength}} \right)$$

While this rule of thumb provides a general trend, it is not a quantitative analysis of LIDT vs wavelength. In CW applications, for instance, damage scales more strongly with absorption in the coating and substrate, which does not necessarily scale well with wavelength. While the above procedure provides a good rule of thumb for LIDT values, please contact Tech Support if your wavelength is different from the specified LIDT wavelength. If your power density is less than the adjusted LIDT of the optic, then the optic should work for your application.

Please note that we have a buffer built in between the specified damage thresholds online and the tests which we have done, which accommodates variation between batches. Upon request, we can provide individual test information and a testing certificate. The damage analysis will be carried out on a similar optic (customer's optic will not be damaged). Testing may result in additional costs or lead times. Contact Tech Support for more information.

Pulsed Lasers

As previously stated, pulsed lasers typically induce a different type of damage to the optic than CW lasers. Pulsed lasers often do not heat the optic enough to damage it; instead, pulsed lasers produce strong electric fields capable of inducing dielectric breakdown in the material. Unfortunately, it can be very difficult to compare the LIDT specification of an optic to your laser. There are multiple regimes in which a pulsed laser can damage an optic and this is based on the laser's pulse length. The highlighted columns in the table below outline the relevant pulse lengths for our specified LIDT values.


Pulses shorter than 10 ns cannot be compared to our specified LIDT values with much reliability. In this ultra-short-pulse regime various mechanics, such as multiphoton-avalanche ionization, take over as the predominate damage mechanism [2]. In contrast, pulses between 10⁻⁷ s and 10⁻⁴ s may cause damage to an optic either because of dielectric breakdown or thermal effects. This means that both CW and pulsed damage thresholds must be compared to the laser beam to determine whether the optic is suitable for your application.

Pulse Duration	t < 10 ⁻⁹ s	10 ⁻⁹ < t < 10 ⁻⁷ s	10 ⁻⁷ < t < 10 ⁻⁴ s	t > 10 ⁻⁴ s
Damage Mechanism	Avalanche Ionization	Dielectric Breakdown	Dielectric Breakdown or Thermal	Thermal
Relevant Damage Specification	N/A	Pulsed	Pulsed and CW	CW

When comparing an LIDT specified for a pulsed laser to your laser, it is essential to know the following:

1. Wavelength of your laser
2. Energy density of your beam (total energy divided by 1/e² area)
3. Pulse length of your laser
4. Pulse repetition frequency (prf) of your laser
5. Beam diameter of your laser (1/e²)
6. Approximate intensity profile of your beam (e.g., Gaussian)

The energy density of your beam should be calculated in terms of J/cm². The graph to the right shows why expressing the LIDT as an energy density provides the best metric for short pulse sources. In this regime, the LIDT given as an energy density can be applied to any beam diameter; one does not need to compute an adjusted LIDT to adjust for changes in spot size. This calculation assumes a uniform beam intensity profile. You must now adjust this energy density to account for hotspots or other nonuniform intensity profiles and roughly calculate a maximum energy density. For reference a Gaussian beam typically has a maximum energy density that is twice that of the 1/e² beam.



LIDT in energy density vs. pulse length and spot size. For short pulses, energy density becomes a constant with spot size. This graph was obtained from [1].

Now compare the maximum energy density to that which is specified as the LIDT for the optic. If the optic was tested at a wavelength other than your operating wavelength, the damage threshold must be scaled appropriately [3]. A good rule of thumb is that the damage threshold has an inverse square root relationship with wavelength such that as you move to shorter wavelengths, the damage threshold decreases (i.e., a LIDT of 1 J/cm² at 1064 nm scales to 0.7 J/cm² at 532 nm):

$$\text{Adjusted LIDT} = \text{LIDT Energy} \sqrt{\frac{\text{Your Wavelength}}{\text{LIDT Wavelength}}}$$

You now have a wavelength-adjusted energy density, which you will use in the following step.

Beam diameter is also important to know when comparing damage thresholds. While the LIDT, when expressed in units of J/cm², scales independently of spot size; large beam sizes are more likely to illuminate a larger number of defects which can lead to greater variances in the LIDT [4]. For data presented here, a <1 mm beam size was used to measure the LIDT. For beams sizes greater than 5 mm, the LIDT (J/cm²) will not scale independently of beam diameter due to the larger size beam exposing more defects.

The pulse length must now be compensated for. The longer the pulse duration, the more energy the optic can handle. For pulse widths between 1 - 100 ns, an approximation is as follows:

$$\text{Adjusted LIDT} = \text{LIDT Energy} \sqrt{\frac{\text{Your Pulse Length}}{\text{LIDT Pulse Length}}}$$

Use this formula to calculate the Adjusted LIDT for an optic based on your pulse length. If your maximum energy density is less than this adjusted LIDT maximum energy density, then the optic should be suitable for your application. Keep in mind that this calculation is only used for pulses between 10⁻⁹ s and 10⁻⁷ s. For pulses between 10⁻⁷ s and 10⁻⁴ s, the CW LIDT must also be checked before deeming the optic appropriate for your application.

Please note that we have a buffer built in between the specified damage thresholds online and the tests which we have done, which accommodates variation between batches. Upon request, we can provide individual test information and a testing certificate. Contact Tech Support for more information.

[1] R. M. Wood, *Optics and Laser Tech.* **29**, 517 (1997).

[2] Roger M. Wood, *Laser-Induced Damage of Optical Materials* (Institute of Physics Publishing, Philadelphia, PA, 2003).

[3] C. W. Carr *et al.*, *Phys. Rev. Lett.* **91**, 127402 (2003).

[4] N. Bloembergen, *Appl. Opt.* **12**, 661 (1973).

LIDT CALCULATIONS

In order to illustrate the process of determining whether a given laser system will damage an optic, a number of example calculations of laser induced damage threshold are given below. For assistance with performing similar calculations, we provide a spreadsheet calculator that can be downloaded by clicking the button to the right. To use the calculator, enter the specified LIDT value of the optic under consideration and the relevant parameters of your laser system in the green boxes. The spreadsheet will then calculate a linear power density for CW and pulsed systems, as well as an energy density value for pulsed systems. These values are used to calculate adjusted, scaled LIDT values for the optics based on accepted scaling laws. This calculator assumes a Gaussian beam profile, so a correction factor must be introduced for other beam shapes (uniform, etc.). The LIDT scaling laws are determined from empirical relationships; their accuracy is not guaranteed. Remember that absorption by optics or coatings can significantly reduce LIDT in some spectral regions. These LIDT values are not valid for ultrashort pulses less than one nanosecond in duration.


[LIDT Calculator](#)

CW Laser Example

Suppose that a CW laser system at 1319 nm produces a 0.5 W Gaussian beam that has a $1/e^2$ diameter of 10 mm. A naive calculation of the average linear power density of this beam would yield a value of 0.5 W/cm, given by the total power divided by the beam diameter:

$$\text{Linear Power Density} = \frac{\text{Power}}{\text{Beam Diameter}}$$

However, the maximum power density of a Gaussian beam is about twice the maximum power density of a uniform beam, as shown in the graph to the right. Therefore, a more accurate determination of the maximum linear power density of the system is 1 W/cm.



A Gaussian beam profile has about twice the maximum intensity of a uniform beam profile.

An AC127-030-C achromatic doublet lens has a specified CW LIDT of 350 W/cm, as tested at 1550 nm. CW damage threshold values typically scale directly with the wavelength of the laser source, so this yields an adjusted LIDT value:

$$\text{Adjusted LIDT} = \text{LIDT Power} \left(\frac{\text{Your Wavelength}}{\text{LIDT Wavelength}} \right)$$

The adjusted LIDT value of 350 W/cm \times (1319 nm / 1550 nm) = 298 W/cm is significantly higher than the calculated maximum linear power density of the laser system, so it would be safe to use this doublet lens for this application.

Pulsed Nanosecond Laser Example: Scaling for Different Pulse Durations

Suppose that a pulsed Nd:YAG laser system is frequency tripled to produce a 10 Hz output, consisting of 2 ns output pulses at 355 nm, each with 1 J of energy, in a Gaussian beam with a 1.9 cm beam diameter ($1/e^2$). The average energy density of each pulse is found by dividing the pulse energy by the beam area:

$$\text{Energy Density} = \frac{\text{Pulse Energy}}{\text{Beam Area}}$$

As described above, the maximum energy density of a Gaussian beam is about twice the average energy density. So, the maximum energy density of this beam is $\sim 0.7 \text{ J/cm}^2$.

The energy density of the beam can be compared to the LIDT values of 1 J/cm² and 3.5 J/cm² for a BB1-E01 broadband dielectric mirror and an NB1-K08 Nd:YAG laser line mirror, respectively. Both of these LIDT values, while measured at 355 nm, were determined with a 10 ns pulsed laser at 10 Hz. Therefore, an adjustment must be applied for the shorter pulse duration of the system under consideration. As described on the previous tab, LIDT values in the nanosecond pulse regime scale with the square root of the laser pulse duration:

$$\text{Adjusted LIDT} = \text{LIDT Energy} \sqrt{\frac{\text{Your Pulse Length}}{\text{LIDT Pulse Length}}}$$

This adjustment factor results in LIDT values of 0.45 J/cm² for the BB1-E01 broadband mirror and 1.6 J/cm² for the Nd:YAG laser line mirror, which are to be compared with the 0.7 J/cm² maximum energy density of the beam. While the broadband mirror would likely be damaged by the laser, the more specialized laser line mirror is appropriate for use with this system.

Pulsed Nanosecond Laser Example: Scaling for Different Wavelengths

Suppose that a pulsed laser system emits 10 ns pulses at 2.5 Hz, each with 100 mJ of energy at 1064 nm in a 16 mm diameter beam (1/e²) that must be attenuated with a neutral density filter. For a Gaussian output, these specifications result in a maximum energy density of 0.1 J/cm². The damage threshold of an NDUV10A Ø25 mm, OD 1.0, reflective neutral density filter is 0.05 J/cm² for 10 ns pulses at 355 nm, while the damage threshold of the similar NE10A absorptive filter is 10 J/cm² for 10 ns pulses at 532 nm. As described on the previous tab, the LIDT value of an optic scales with the square root of the wavelength in the nanosecond pulse regime:

$$\text{Adjusted LIDT} = \text{LIDT Energy} \sqrt{\frac{\text{Your Wavelength}}{\text{LIDT Wavelength}}}$$

This scaling gives adjusted LIDT values of 0.08 J/cm² for the reflective filter and 14 J/cm² for the absorptive filter. In this case, the absorptive filter is the best choice in order to avoid optical damage.

Pulsed Microsecond Laser Example

Consider a laser system that produces 1 µs pulses, each containing 150 µJ of energy at a repetition rate of 50 kHz, resulting in a relatively high duty cycle of 5%. This system falls somewhere between the regimes of CW and pulsed laser induced damage, and could potentially damage an optic by mechanisms associated with either regime. As a result, both CW and pulsed LIDT values must be compared to the properties of the laser system to ensure safe operation.

If this relatively long-pulse laser emits a Gaussian 12.7 mm diameter beam (1/e²) at 980 nm, then the resulting output has a linear power density of 5.9 W/cm and an energy density of 1.2 x 10⁻⁴ J/cm² per pulse. This can be compared to the LIDT values for a WPQ10E-980 polymer zero-order quarter-wave plate, which are 5 W/cm for CW radiation at 810 nm and 5 J/cm² for a 10 ns pulse at 810 nm. As before, the CW LIDT of the optic scales linearly with the laser wavelength, resulting in an adjusted CW value of 6 W/cm at 980 nm. On the other hand, the pulsed LIDT scales with the square root of the laser wavelength and the square root of the pulse duration, resulting in an adjusted value of 55 J/cm² for a 1 µs pulse at 980 nm. The pulsed LIDT of the optic is significantly greater than the energy density of the laser pulse, so individual pulses will not damage the wave plate. However, the large average linear power density of the laser system may cause thermal damage to the optic, much like a high-power CW beam.

Part Number	Description	Price	Availability
CCM5-F01/M	Customer Inspired!16 mm Cage-Cube-Mounted UV Enhanced Aluminum Turning Prism Mirror, M4 Tap	\$130.00	Today
CCM5-G01/M	Customer Inspired!16 mm Cage-Cube-Mounted Protected Aluminum Turning Prism Mirror, M4 Tap	\$130.00	Today
CCM5-P01/M	Customer Inspired!16 mm Cage-Cube-Mounted Silver Turning Prism Mirror, M4 Tap	\$130.00	Today
CCM5-M01/M	Customer Inspired!16 mm Cage-Cube-Mounted Gold Turning Prism Mirror, M4 Tap	\$130.00	Today
CCM5-E02/M	Customer Inspired!16 mm Cage-Cube-Mounted Dielectric Turning Prism Mirror, 400 - 750 nm, M4 Tap	\$200.00	Today
CCM5-E03/M	Customer Inspired!16 mm Cage-Cube-Mounted Dielectric Turning Prism Mirror, 750 - 1100 nm, M4 Tap	\$205.00	Today
CCM5-K13/M	Customer Inspired!16 mm Cage-Cube-Mounted Nd:YAG Turning Prism Mirror, 532 and 1064 nm, M4 Tap	\$235.00	Today
CM05-F01	16 mm Cage-Cube-Mounted UV Enhanced Aluminum Turning Prism Mirror, 8-32 and M4 Adapters	\$134.00	Today
CM05-G01	16 mm Cage-Cube-Mounted Protected Aluminum Turning Prism Mirror, 8-32 and M4 Adapters	\$134.00	Today

CM05-P01	16 mm Cage-Cube-Mounted Silver Turning Prism Mirror, 8-32 and M4 Adapters	\$134.00	Today
CM05-M01	16 mm Cage-Cube-Mounted Gold Turning Prism Mirror, 8-32 and M4 Adapters	\$134.00	Today
CM05-E02	16 mm Cage-Cube-Mounted Dielectric Turning Prism Mirror, 400 - 750 nm, 8-32 and M4 Adapters	\$206.00	Today
CM05-E03	16 mm Cage-Cube-Mounted Dielectric Turning Prism Mirror, 750 - 1100 nm, 8-32 and M4 Adapters	\$211.00	Today
CM05-K13	16 mm Cage-Cube-Mounted Nd:YAG Turning Prism Mirror, 532 and 1064 nm, 8-32 and M4 Adapters	\$242.00	Today

Visit the *16 mm Cage Cube-Mounted Turning Prism Mirrors* page for pricing and availability information:

https://www.thorlabs.com/newgrouppage9.cfm?objectgroup_id=5049